



Assembly Instructions

INFM7X05, INFM7X06, INFM7X09,
INFM7X10, INFM7X11, INFM7X16,
INFM7X20, INFM7X20L, INFM7X22,
INFM7X22L, INFM7X23, INFM7X25,
INFM7X25L, INFM7X28, INFM7X30,
INFM7X30L, INFM7X32, INFM7X35,
INFM7X40

Infinity Mounting set
Universal

1. Safety

Symbols



Warning!

Risk of serious injury or death.

An operation that, if not correctly performed, may lead to injury to the operator or others.



CAUTION!

Risk of material damage.

An operation that, if not correctly performed, may lead to material damage.



INFO

Particularly important or explanatory information

General instructions

These instructions shall always be followed when placing the mounting set INFM7 Universal on a wheelchair.

To be able to connect Infinity to the wheelchair Infinity also needs to be adjusted according to the assembly instructions that is included in the delivery of Infinity.


Warnings



WARNING!

Assembly work or changes made to the mounting set must be carried out by trained service technician, reseller or other person assigned by Decon.

2. Overview



A Center pipe adapter bar

B Outer pipe adapter bar

C Clamp adapter bar

D Holder adapter bar

E QR-Pin

F Bar

G Chassis clamp

H Holder screw



The shape of the chassis clamp (G) can vary depending on the tube size of the chassis and might not match the clamps in the picture above.

3. Assembly and adjustment of adaptor

First mount the holder (D) onto the bar (F) and secure it by using the screws (H). Now mount the preassembled parts onto the wheelchair by using the chassis clamps (G), see picture as example. The distance **X** from the edge of the seating cushion to the holder (D) should be, if possible, **20-30mm** so that the steering plates of the center pipe adapter bar (A) is visible outside the edge of the seating cushion.

The distance **Y** should not be greater than **20mm** to ensure that the bar (F) gets sufficient support. The clamp (G) should be mounted as close as possible to the holder (D).

Tighten the screws in the clamps (G) just enough to grip the wheelchair chassis so that the bar (F) can still be adjusted inside the clamps.

Assemble both sides.



INFO

In some cases the chassis clamp (G) has to be rotated slightly outward or inwards to make space for the original brakes of the wheelchair.




INFO

In some cases, it can be preferable to remove the original brakes of the wheelchair to make room for assembling of the adaptor bar. On some chairs it might also be needed to loosen the seat cloth a little bit to make room for the chassis clamps (G).



WARNING!


The screws (H) must be tightened to 10-15Nm of torque. Loctite 243 or eq. shall be applied to the screw before it is assembled.






WARNING!

One clamp (G) is to be mounted on each side of the holder (D) to avoid that the bar breaks due to fatigue stress.



Now mount the complete adapter bar that consists of center pipe (A), outer pipe (B) and clamp (C). The complete adapter bar is mounted into the holders (D) by using the QR-pins (E).

Make sure the ends of the outer pipe (B) are parallel otherwise it can be hard to assemble it into the holders (D). It shall be possible to move the center pipe sideways and also rotate it if correctly assembled.



Now tighten the screws in all of the clamps (G).



INFO

Check that the arrow on the labeling mounted on the center pipe (A) is pointing towards the front of the wheelchair. If the center pipe (A) is mounted the wrong way around it will not be possible to attach Infinity to the bar.




WARNING!

The screws holding the clamps (G) shall be tightened to:
7-9Nm of torque (M6)
10-15Nm of torque (M8)

Now use the leveling tool that is included in the delivery of Infinity. Since the adapter bar have splines inside it will not be possible to get the bubble of the spirit level exactly right. Just choose the most correct position.


The leveling tool has two marks “H” and “L”. When the bubble of the spirit level is towards the “H” the castors of the wheelchair will lift higher from the ground compared to if the bubble of the spirit level is towards the “L”.



Adjustment of the angle can be done in two different manners. Both is described bellow.


Alt1.

Using the adjustment tool that is included in the delivery of Infinity (not this mounting set), set the angle of the adaptorbar so that the spirit level is in level. Inside the pipes there are splines so when rotating you will feel and hear a clicking. When desired angle have been accomplished, check that the center pipe is centered in the wheelchair. Now tighten the screws in the clamps (C).




WARNING!

The screws holding the clamps (C) shall be tightened to 7-9Nm of torque.




Alt.2

Using the adjustment tool that is included in the delivery of Infinity (not this mounting set), set the angle of the adaptorbar. Only mount the inner pipe on one side and then rotate the center pipe to the correct angle. Now mount the other inner pipe in the same position/angle as the first one. Check that the center pipe is centered in the wheelchair. Now tighten the screws in the clamps (C).



WARNING!

The screws holding the clamps (C) shall be tightened to 7-9Nm of torque.



Now read the assembly instructions that was included in the delivery of the Infinity unit. Assemble and adjust the Infinity according to the instruction.



INFO

Don't forget to reassemble the brakes of the wheelchair in case these were removed when mounting the adaptor. Also, if the seat cloth was loosened during mounting of the adaptor don't forget to reassemble this.

Decon Wheel AB
Södra Ekeryd 119, SE-314 93 Hyltebruk, Sweden
Tel: +46 (0)345 40880
Fax: +46 (0)345 40895
E-mail: info@decon.se

